

Reagent Disks Specifications

Freshwater FF104 (Code 4353)

Test Factor	Display Abbreviation	Range	Accuracy	Method Detection Limit
Alkalinity	ALK/D ALK	0 – 250 ppm/14.0 dKH	± 15%	15 ppm/0.8 dKH
Ammonia	AMMO	0.0 – 3.0 ppm	< 2.0 ppm: ± 0.2 ppm > 2.0 ppm: ± 0.4 ppm	0.2 ppm
Hardness	G HARD/D HARD	0 – 500 ppm/ 28.0 dGH	± 15%	20 ppm/1.1 dGH
Nitrate	NITRATE	0 – 300 ppm	± 20%	5 ppm
Nitrite	NITRITE	0.0 – 2.0 ppm	± 0.2 ppm	0.1 ppm
pH	pH	4.5 – 10.0	± 0.2	NA
Phosphate	PHOS	0.0 – 2.0 ppm	± 0.2 ppm	0.2 ppm

Saltwater FF203 (Code 4354)

Test Factor	Display Abbreviation	Range	Accuracy	Method Detection Limit
Alkalinity	ALK/ALK D	0 – 300 ppm/17.0 dKH	± 15%	15 ppm/0.8 dKH
Ammonia	AMMO	0.0 – 3.0 ppm	< 1.0 ppm: ± 0.2 ppm > 1.0 ppm: ± 0.4 ppm	0.2 ppm
Calcium	Ca	200 – 800 ppm	± 15%	NA
Magnesium	Mg	500 – 2200 ppm	± 15%	NA
Nitrate	NITRATE	0 – 60 ppm	± 20%	5 ppm
Nitrite	NITRITE	0.0 – 2.0 ppm	± 0.2 ppm	0.1 ppm
pH	pH	6.5 – 10.0	± 0.2	NA
Phosphate	PHOS	0.0 – 2.0 ppm	± 0.2 ppm	0.2 ppm

